

Declaration of Conformity to Council Directive 93/42 EEC (Including Directive 2007/47/EEC) Concerning Medical Devices

Manufacturer: Shenzhen Luckcome Technology Inc., Ltd.
201, 2F, NO.1 ZhongJian Industrial Building, NO.18
Yanshan Road, SheKou, Nanshan District,
Shenzhen. Guangdong, China

European Representative: MEQUIPEX
Feldstrasse 39, 4813 Altmuenster Austria

Product: Fetal Doppler
MODEL: L6、L6C、L6S、L6SE、L6T、FD88

Classification: II a (Rule 10 of Annex IX, MDD)

Conformity assessment route: Annex II excluding (4)

We, the manufacturer, herewith declare under our sole responsibility that the stated medical devices meet the transposition into national law, the provisions of Council Directive 93/42/EEC of 14 June, 1993, concerning medical devices; including the amendments by Council Directive 2007/47/EEC. All supporting documentation is retained at the premises of the manufacturer.

The validity period of this declaration of conformity is limited by the issuing of a revised declaration of conformity after change of the product and/or by the expiration date of the related Annex II certificate issued by the notified body.

Standards applied: Applied Standards List (attached) for which documented evidence of compliance can be provided.

Notified Body: MEDCERT GmbH
Pilatuspool 2, D-20355 Hamburg, Germany

Identification Number:  0482

(EC) Certificate(s): 6016GB410141217

Start of CE-marking: 17 December, 2014

Place, Date of Issue: 5 June, 2016

Signature: Name: Huang Ping
Position: General Manager

List of Harmonized standard

| No. | Standard Name | Reference No. |
|-----|--|------------------------------------|
| 1 | Medical device risk management to medical devices application | EN ISO 14971:2012 |
| 2 | Symbol for the label of medical devices | EN ISO 15223-1:2012 |
| 3 | Term, symbol and information of medical device— information of medical device manufacturer offering | EN 1041:2008 |
| 4 | Medical electrical equipment — Part 1: General requirements for basic safety and essential performance | EN 60601-1:2006 / AC:2010 |
| 5 | Medical electrical equipment - Part 1-2: General requirements for safety - Collateral standard: Electromagnetic compatibility - Requirements and tests | EN 60601-1-2:2007 |
| 6 | Medical electrical equipment -- Part 1-6: General requirements for basic safety and essential performance - Collateral standard: Usability | EN 60601-1-6:2010 |
| 7 | Medical electrical equipment Part 2-37: Particular requirements for the safety of ultrasonic medical diagnostic and monitoring equipment | EN 60601-2-37:2008 |
| 8 | Clinical thermometers - Part 4: Performance of electrical thermometers for continuous measurement | EN 12470-4:2000+A1:2009 |
| 9 | Non-invasive sphygmomanometers - Part 3: Supplementary requirements for electro-mechanical blood pressure measuring systems | EN 1060-3:1997+A2:2009 |
| 10 | Biological evaluation of medical devices - Part 1: Evaluation and testing within a risk management process | EN ISO 10993-1:2009 |
| 11 | Biological evaluation of medical devices - Part 5: Tests for in vitro cytotoxicity | EN ISO 10993-5:2009 |
| 12 | Biological evaluation of medical devices Part 10: Tests for irritation and skin sensitization | EN ISO 10993-10:2010 |
| 13 | Medical device software - Software life-cycle processes | EN 62304:2006 |
| 14 | Packaging-Pictorial marking for handling of goods | ISO 780 :1997 |
| 15 | Sterilization of medical devices Information to be provided by the manufacturer for the processing of resterilizable medical devices | ISO 17664:2004 |
| 16 | Standard Test Method to Determine Efficacy of Disinfection Processes for Reusable Medical Devices (Simulated Use Test) | ASTM E 1837 – 96 (Reapproved 2002) |
| 17 | Non-invasive sphygmomanometers - Part 4: Test procedures to determine the overall system accuracy of automated non-invasive sphygmomanometers | EN 1060-4:2004 |