



## EU Quality Management System Certificate (MDR)

Pursuant to Regulation (EU) 2017/745 on Medical Devices, Annex IX Chapters I and III  
(Class I Devices in sterile condition, with measuring function or reusable surgical instruments)

**No. G11 075321 0012 Rev. 00**

**Manufacturer:** **Guangdong Kindly Medical Device Group Co., Ltd.**  
Room101, Building A  
No.288 East Airport Road, Sanzao Town, Jinwan District  
519040 Zhuhai, Guangdong  
PEOPLE'S REPUBLIC OF CHINA

**SRN Manufacturer:** CN-MF-000006505

**Authorized Representative:** Shanghai International Holding Corp. GmbH (Europe)  
Eiffestraße 80, 20537 Hamburg, GERMANY

The Certification Body of TÜV SÜD Product Service GmbH certifies that the manufacturer has established, documented and implemented a quality management system as described in Article 10 (9) of the Regulation (EU) 2017/745 on medical devices. Details on device categories covered by the quality management system are described on the following page(s). The Report referenced below summarises the result of the assessment and includes reference to relevant CS, harmonized standards and test reports. The conformity assessment has been carried out according to Annex IX Chapter I and III of this regulation with a positive result. As applicable the involvement of the notified body is limited to the aspects relating to:

- establishing, securing and maintaining sterile conditions,
- conformity of the devices with the metrological requirements,
- reuse of the device, in particular cleaning, disinfection, sterilization, maintenance and functional testing and the related instructions for use.

The certified quality management system is subject to periodical surveillance by TÜV SÜD Product Service GmbH. All applicable requirements of the testing and certification regulation of TÜV SÜD Group have to be complied with.

For details and certificate validity see: [www.tuvsud.com/ps-cert?q=cert:G11 075321 0012 Rev. 00](http://www.tuvsud.com/ps-cert?q=cert:G11_075321_0012_Rev.00)

**Report No.:** BJ21077201

**Valid from:** 2023-02-14

**Valid until:** 2028-02-13

Christoph Dicks  
Head of Certification/Notified Body

**Issue date:** 2023-02-14



## EU Quality Management System Certificate (MDR)

Pursuant to Regulation (EU) 2017/745 on Medical Devices, Annex IX Chapters I and III  
 (Class I Devices in sterile condition, with measuring function or reusable surgical instruments)

**No. G11 075321 0012 Rev. 00**

**Classification:** I  
**Device Group:** A020102 - INFUSION AND IRRIGATION SYRINGES, SINGLE-USE  
**Device Properties:** MDS 1005.1 - Ethylene Oxide sterilization  
 MDS 1010 - Devices with a measuring function

**The validity of this certificate depends on conditions and/or is limited to the following:** not yet available