



Friedrich Bosch GmbH & Co. KG · Postfach 5 · D-72402 Bisingen

Hohenlalienstraße 30
D-72406 Bisingen
Telefon +49 7476 3554, 3555
Telefax +49 7476 3556
info@fbosch.de
www.fbosch.de

DECLARATION OF CONFORMITY

We company **FRIEDRICH BOSCH GMBH&CO.KG, Hohenlalienstr. 30, 72406 Bisingen, Germany**

declare under our sole responsibility that the product:

aneroid sphygmomanometer Diagnostic Set

articlenumber:	9900.99.112	black
	9900.99.105	burgundy
	9900.99.108	green
	9900.99.109	yellow
	9900.99.110	orange
	9900.99.116	red
	9900.99.117	blue

class of product: class **I M**

to which this declaration relates is in conformity with the following standards
Guideline 93/42 EWG appendix VI.3

The quality system is attested by the accredited body TÜV Süd PRODUCT SERVICE GmbH, Ridlerstr. 65, 80339 München with the licence number **0123**.

The mentioned product is manufactured in compliance with EN 1060-1 and 2: 1994.

The certificate has the number **G3M 020267 0020 Rev. 00** and is valid until **26.05.2024**.

date: 09.05.2021

Josef Bieger
Director of quality-management

Friedrich Bosch GmbH & Co. KG
MEDIZINTECHNIK
Postfach 5 | Hohenlalienstr. 30
D-72402 Bisingen | D-72406 Bisingen
Tel. 07476/3554, 3555 · Fax 07476/3556